



## **EOI SCHEDULE - I**

**FOR SUPPLY, INSTALLATION, COMMISSIONING,  
VALIDATION OF EQUIPMENT AT SOIL TESTING  
LABS OF DEPARTMENT OF AGRICULTURE,  
ANDHRA PRADESH 2018-19**



**ANDHRA PRADESH STATE AGRO INDUSTRIES DEVELOPMENT CORPORATION LIMITED**  
#55-17-2 to 4, 5<sup>th</sup> Floor, Stalin Corporate, Industrial Estate, Autonagar, Vijayawada – 520007,  
Ph.No.0866-2556803, FAX:0866-2556804  
Email: [apagros@ap.nic.in](mailto:apagros@ap.nic.in), web site: [www.apagros.org](http://www.apagros.org)



**ANDHRA PRADESH STATE**  
**AGRO INDUSTRIES DEVELOPMENT CORPORATION LTD**

**(A GOVT. UNDERTAKING)**

#55-17-2 to 4, 5<sup>th</sup> Floor, Stalin Corporate, Industrial Estate, Autonagar, Vijayawada – 520007,  
Ph.No.0866-2556803, FAX:0866-2556804

[E.Mail: apagrosdmae@gmail.com](mailto:apagrosdmae@gmail.com)

**No. Agros/FMD/EOI/ST labs/2018-19**

**Dt:20-06-2019**

**SHORT EOI NOTICE INVITING SEALED QUOTATIONS**

Sealed Quotations are invited from reputed manufacturers, Distributors, Dealers & Importers for the following items:

Description of Item	Cost of EOI schedule	Sale of EOI Schedules From – to	Last date for submission	Date of opening of Technical Quotations
For Supply, Installation commissioning, Validation of Soil Testing Lab Equipment	Rs.10,000	24-06-2019 to 04-07-2019 up to 2.00 P.M.	04 -07-2019 up to 2.30 P.M.	04-07-2019 at 03.30 P.M.
For Supply, Installation commissioning, Validation of Pesticide Testing Lab Equipment	Rs.10,000	24-06-2019 to 05-07-2019 up to 2.00 P.M.	05 -07-2019 up to 2.30 P.M.	05-07-2019 at 03.30 P.M.

The EOI Schedules and terms and conditions can be downloaded from our website [www.apagros.org](http://www.apagros.org)

Vijayawada

Sd/-  
Vice Chairman and  
Managing Director



**ANDHRA PRADESH STATE**  
**AGRO INDUSTRIES DEVELOPMENT CORPORATION LTD**

**(A GOVT. UNDERTAKING)**

#55-17-2 to 4, 5<sup>th</sup> Floor, Stalin Corporate, Industrial Estate, Autonagar, Vijayawada – 520007,  
Ph.No.0866-2556803, FAX:0866-2556804

[E.Mail: apagrosdmc@gmail.com](mailto:apagrosdmc@gmail.com)

**EOI SCHEDULE**

**Name of work:** FOR SUPPLY, INSTALLATION, COMMISSIONING, VALIDATION OF  
EQUIPMENT AT SOIL TESTING LABS OF DEPARTMENT OF AGRICULTURE, ANDHRA  
PRADESH 2018-19

Ref: AGROS/FMD /EOI/LAB EQUIPMENT/2019-20, Dated : 20 -06-2019

**INSTRUCTIONS TO OFFERER(S)**

**A. PREAMBLE :**

The Department of Agriculture, Government of Andhra Pradesh are requiring various types of Lab equipment., to their Soil Testing Labs in the State of Andhra Pradesh through the Nodal Agency A.P.S Agro Ind. Dev. Corp. Ltd. (APSAIDCL), a Government of Andhra Pradesh undertaking for procurement and commissioning of Lab equipment.

**B. ELIGIBLE CONDITIONS FOR EOI SCHEDULE:**

1. The manufacturers / importers/distributors should have minimum of 3 years experience in supply and commissioning of scientific equipment to the Government Departments / Institutes, such as Universities. Firms must furnish PO copies/ installation reports/ **performance certificate** from concerned Department / Institution in support of the above.
2. The firm must have works contract registration and GST registration certificate from appropriate authorities (copies to be enclosed.)
3. The firms shall have GST (TIN) Registration and PAN Number.
4. The firms shall have annual turnover of not less than Rupees one crore in last 3 years 2016-17,17-18 and 2018-19 and proof of same to be submitted.
5. The firm who are importing equipment must have importing license.
6. The Distributors who are participating in the EOIs should enclose authorization letter from the Principal manufacturer.
7. The cost of EOI schedule of Rs.10000/- by way of demand draft drawn in favor of APSAIDC.Ltd Vijayawada shall be enclosed.
8. Number of technicians / workers list to be produced with their qualification.

### **C) INSTRUCTIONS FOR SUBMISSION OF EOI SCHEDULE**

1. The Offerer shall submit filled in EOI Schedule with all documents along with the cost EOI schedule , covering letter with EMD in a separate sealed cover duly superscribing “Technical quotation for EOI Schedule for Supply, Installation, Commissioning & Validation of lab equipment” and Price Quote as per Annexure-III in a separate sealed cover duly superscribing “Price quote for Supply, Installation, Commissioning & Validation of lab equipment” with name and address of the offerer. Both covers are to be placed in a third cover duly superscribing “ EOI Schedule for Supply, Installation, Commissioning & Validation of lab equipment.”
2. The offerer has to deposit an amount of Rs:2,00,000/- for the items indicated at SI No.1 to 8 and Rs.50000/- for the items SI No.9 to 11 towards EMD by way of demand draft drawn in favor of **A.P.S.A.I.D.C. Ltd., payable at Vijayawada** along with a covering letter. EOI Schedules without EMD will be rejected.
3. The Earnest Money Deposit of successful firm shall be retained by the Corporation and refunded only after satisfactory completion of the contract. The EMD can be adjusted against performance guarantee deposit for successful offerer(s). The E.M.D. and performance guarantee deposit will not carry any interest.
4. Performance guarantee deposit will be maximum Rs.2,00,000/- and minimum Rs.50,000/- for the equipment @ **5% on the total value of equipment** shall be paid in favor of A.P.S.A.I.D.C. Ltd., payable at Vijayawada in the shape of Demand Draft drawn from any Indian Nationalized Bank. The performance guarantee shall be valid for 1 year.
5. a) Performance guarantee produced by the successful offerer shall be forfeited and alternate arrangements shall be made, if the firm does not fulfill the obligation of supply of stocks as per orders / specification /quality and alternate arrangements will be made for procurement of such products and any additional cost incurred in this regard will be recovered from the concerned firm.  
b) The responsibility of quality and quantity of the lab equipment supplied lies with supplier only.
6. The Offerer shall submit duly notarized declaration along with Technical bid, as per the proforma furnished under Annexure-IV, hereto on Rs.100/- non-judicial stamp paper.
7. The authorized person representing the firm who is signing in the EOI documents and also participating in the EOI opening shall invariably attach a copy of authorization from the competent authority of the firm as per the By-laws of the firm / Company.
8. The last date for submission of filled in EOI schedule is **upto 02.30 P.M. on 04-07-2019**, the technical bid will be opened at 03.30 P.M. on the same day.

### **D) GENERAL TERMS & CONDITIONS**

1. The price quote of only those offerer(s) whose Technical quotation are qualified will be opened with due intimation to them.

2. Negotiations will be held as per the decision of the Technical Committee at the address indicated in the data sheet. The aim is to reach agreement on all the points and sign a contract.
3. The **agreement period will be one year from the date of signing on the agreement**. No escalation charges will be paid for supply of items during the agreement period. The validity period of the final negotiated prices can be extended on mutual consent of the both the parties.
4. The supplies made shall be invoiced on GST invoice with Serial No. of invoice printed on it. No invoice other than tax invoice will be accepted. GST applicable, if any, shall be clearly mentioned in the tax invoice.
5. The approved firm shall supply the ordered equipment as per terms and conditions within the stipulated time indicated in the Purchase Order.
6. In the case of Annexure-II, the items supplied must be covered with warranty for minimum of one year (or more as per manufacturer's terms & conditions) from the date of supply against any defective manufacturing / workmanship etc. Repairs / replacements / servicing shall be attended at free of cost within the warranty period.
7. The firm should submit the Utilization/ Installation Certificate and delivery challan in token of receipt of equipment from the Assistant Director of Agriculture of concerned lab for effecting the payment.
8. The loading and unloading charges, if any, should be borne by the supplier only.
9. The firm shall provide detailed operational and maintenance manuals at the time of supply of the equipment.
10. The offerer has to quote the price only as per the technical specifications given in the EOI schedule.
11. Andhra Pradesh State Agro Industries Development Corporation Limited (APSAIDC Ltd) reserves the right to inspect the manufacturing / other premises of supplier and his principal/s during the EOI process or before supply or before payment in order to satisfy that the supply is as per the quality and specifications of equipment indicated in this EOI.
12. Andhra Pradesh State Agro Industries Development Corporation Limited reserves the right to award the EOI either in part or full or follow any other suitable procurement procedure. AP AGROS at its sole discretion and without assigning any reason thereof, reserves the right to accept / reject any or all schedules at any stage of the EOI process.
13. In case of any dispute or difference arising out of the Agreement between the two parties relating to any aspect of the agreement, the parties should attempt to settle the dispute through mutual and amicable discussions. If the dispute is not settled through such discussions, the matter shall be referred / settled as per provisions of Arbitration and Conciliation Act, 1996.

14. It is further to be agreed by both the parties that all differences and disputes if not resolved as above, shall be got remedied through the Court of Law of Amravati Guntur jurisdiction only.
15. EOI can be rejected at any stage of the procurement process, it comes to knowledge that any offerer is involved in any fraudulent practice or misrepresentation of the facts or tried to influence the selection process or indulged in any corrupt activities in competing with others.
16. The manufacturer/ importer/distributor shall be liable for blacklisting and further prosecution, if they fail to adhere to the above conditions.
17. The approved firm shall enter into an agreement with the Corporation duly incorporating all the above terms & conditions.
18. The equipment shall be supplied & commissioned in the Soil Testing Laboratories of Department of Agriculture, Andhra Pradesh at required destinations of Andhra Pradesh State. The offerer shall provide all information as per this EOI document. AP AGROS shall evaluate only those schedules that are received in the required format and are complete in all respects.
19. L1 Price will be taken as bench mark or any suitable procedure will be followed as per discretion of the EOI committee to finalize the prices of Lab equipment.

**E) INSTALLATION, COMMISSIONING & VALIDATION**

1. All tools, equipment and machinery required for the installation and commissioning work has to be arranged by the successful firm only.
2. The period stipulated for completion of supply, installation, commissioning, validation and training of personnel shall be 20 days from the date of receipt of Purchase Order.
3. Penalty for delay in work will be levied @ 0.5% of the total value of Work per week upto a minimum of 5% in case of default in work completion within the stipulated time.

**F) INSURANCE**

The contractor shall insure all goods for all risks to the extent of value of goods specially mentioning that the claim amount under the policy shall be payable to AP AGROS. The contractor shall borne the insurance charges for all risks for the goods for the period upto expiry of warranty.

**G) WARRANTY**

The performance guarantee period shall be 1 Year from the actual date of supply / commissioning of equipment in all respects. The same shall have to be communicated in writing to AP State Agro Industries Development Corporation Limited based on the actual date of completion of all works.

**H) CERTIFICATION PROCESS FOR PAYMENT**

**a. EQUIPMENT / MATERIAL RECEIPT CERTIFICATE**

Issue of material receipt note / goods receipt notes at site by Agricultural Officer and Assistant Director of Agriculture of concerned Laboratory in charge stating that all the Equipment is received as per the specification. (List to be enclosed)

**b. INSTALLATION & COMMISSIONING CERTIFICATE**

Issue of installation and commissioning Certificate by Assistant Director of Agriculture of concerned Laboratory Incharge .

**c. COMPLETION CERTIFICATE**

Completion Certificate will be issued by A.D.A. of Laboratory Incharge after complete validation and training to the personnel of Agriculture Department.

**I) TERMS OF PAYMENT**

Payment shall be released as indicated below, subject to receipt of funds from Commissioner & Director of Agriculture.

- a) The payment will be made against the submission of bills and delivery challans by the supplying firm duly signed and certified by the Assistant Director of Agriculture incharge of the concerned Laboratory stating that all the equipment is received, installed and commissioned as per the specifications and the work done by the firm is satisfactory.

**J) SPECIAL INSTRUCTIONS TO THE FIRMS**

1. The firms should submit technical specifications / brochures / information, users list etc. along with offer.
2. The supplier should conduct working demonstration of the equipment and train the working personnel of Department of Agriculture of concerned Lab.
3. List of fast moving spares and its selling price should also be enclosed. List of accessories which are recommended to be purchased separately for each item should be clearly mentioned in the offer.
4. The successful supplier will make a pre-delivery presentation about the details of equipment being supplied both to the Officers of AP AGROS and the A.D.A's incharge of concerned laboratory.

Place:  
Date :

SIGNATURE OF THE OFFERER  
WITH SEAL

**SCHEDULE**  
**CONDITIONS OF ELIGIBILITY FOR APPLICANT**  
**(Attested copies wherever relevant to be enclosed)**

1	EOI Reference	:	No:Agros/FMD/EOI/LAB EQUI/2018-19 Date: 20 -06-2019
2	a) Name of the Manufacturing/ supplying Company with its Regn. No. (copy of the manufacturing license to be enclosed) b) Name of the Importing company with its Regn. No. (copy of the Importing license to be enclosed)	:	
3	Status of Company		
	i) Registered Company	:	
	ii) Sole proprietorship firm (Name & address of the proprietor)	:	
	iii) Partnership firm (Partnership agreement, letter of consent as authorized by other partners to attend for this EOI )	:	
	iv) Private Limited Company (Certificate of incorporation and authorization from CEO if not represented by CEO)	:	
4	Registered Address of manufacturer/supplier/Dealer	:	
5	Particulars of Demand draft enclosed towards cost of EOI schedule if the EOI schedule is downloaded from website		DD No. DD Date: for Rs.10,000/- Name of the Bank
6	Details of EMD paid DD No: Date: Bank Name: Amount:	:	
7	Brief Profile of the firm	:	



8	Import License No. (copy to be enclosed)	:	
9	Bank particulars for online transfer of amounts <ul style="list-style-type: none"> <li>• Name of the Bank:</li> <li>• Branch of the Bank</li> <li>• Account Number</li> <li>• RTGS No</li> </ul>	:	
10.	Specific authorization letter from the Principal manufacturing company (if yes copy to be enclosed)	:	
11	Whether the Company is ISO certified or not (If 'yes' copy to be enclosed)	:	
12	Proof of having sufficient no. of experienced technical teams for training & operations management (Details: names of persons, qualifications and experience)	:	
13	Annual production capacity certificate / Annual import capacity certificate	:	
14	Turnover of the above machinery during last 3 years. <p style="text-align: center;">a.2016-17</p> <p style="text-align: center;">b.2017-18</p> <p style="text-align: center;">c.2018-19</p>	:	
15	Experience / performance certificate from clients / Govt. departments/ Organizations during last 3 years about supply of lab equipment, performance reports , PO copies , installation certificate from Govt. Institutes, if any.	:	
16	Details of dealers' net work and after sales service stations to be provided by firm in the area of operation.	:	
17	GST Regn.No. / (copies to be enclosed)	:	
18	PAN No. (copy to be enclosed)	:	
19	Audited accounts balance sheet for the last 3 years. (attested copies to be enclosed)	:	
20	Income Tax assessment Certificate for the last 3 years (attested copies to be enclosed)	:	
21	Name of contact person with address & Phone no. Email:	:	

22	a) Detailed Technical specifications of all the equipment quoted in the Price bid. (List to be enclosed) as per annexure-II	:	
23	B.I.S. Certificate of principal manufacturer (copies to be enclosed)		

**(Add additional sheets as required)**

**NOTE:**

1. All the columns of the application form shall be duly, properly and exhaustively filled in.
2. All corrections if any must be signed by the Offerer before submission itself. For filing faulty information, the EOI is liable for rejection. The Offerer should sign on each and every page of EOI schedule and on all the additional documents submitted along with EOI schedule.
3. The applicant can enclose additional information, if any, relevant to the application separately
4. At the time of verification of EOI schedules, the original documents should be produced for verification.

**SIGNATURE OF OFFERER WITH SEAL**

# **TECHNICAL QUOTATION**

## ANNEXURE- II

### List of Equipment required for supply and commissioning along with Specifications

Sl. No.	Description of equipment	Total quantity required (in Nos.)	Plz indicate tick mark for the items quoted
1	PH METER ELECTRODES	49	
2	EC METER	17	
3	EC METER ELECTRODES	23	
4	AUTO SAMPLER TEST TUBES FOR MPAES-4200	240	
5	ACETYLENE GAS CYLINDER WITH REGULATER	47	
6	UV VISABLE SPECTRAPHOTOMETER WITH PC ATTACHMENT	1	
7	ICPOES(INDUCTIVELY COUPLED PLASMA OPTICAL EMISSION SPECTOMETRY)	11	
8	PH/EC SOIL ANAYLISER (SP 2000 SERIES)	17	
9	SET OF SOLAR BATTERIES (12V/200 AHS TUBLER BATTERIES) WITH 4 YEARS WARRENTY	14 STL(INCLUDING RSTL BAPATLA)	
10	SOLAR PANELS 1 no/ 250 WTS CAPACITY		
11	INVERTERS 5KVA		

## Specifications of the Lab equipment

### **1.PH METER ELECTRODES:**

**( The offerer has to furnish the specifications of the equipment)**

### **2.EC METER:**

**( The offerer has to furnish the specifications of the equipment)**

### **3.EC METER ELECTRODES:**

**( The offerer has to furnish the specifications of the equipment)**

### **4.AUTO SAMPLER TEST TUBES FOR MPAES-4200:**

**( The offerer has to furnish the specifications of the equipment)**

### **5.ACETYLENE GAS CYLINDER WITH REGULATER:**

**( The offerer has to furnish the specifications of the equipment)**

### **6.UV VISABLE SPECTRAPHOTOMETER WITH PC ATTACHMENT:**

**( The offerer has to furnish the specifications of the equipment)**

## 7. Specifications of true Simultaneous ICP-AES/OES:

<b>Scope of Analysis</b>	ICP-OES/AES system for analysis of soil and water.	
<b>Spectrometer</b>	Poly chromatic design True Simultaneous spectrometer with full PC control system.	
<b>Wavelength Range</b>	170 to 800 nm or better	
<b>Resolution</b>	Should be < 0.005 nm @ 200 nm or better	
<b>Optical system</b>	Optical system should be echelle based grating with min 79 lines/mm. The Reciprocal linear desperation should be both in UV and VIS region.	
<b>Viewing Angle</b>	Both axial and radial viewing within single system.	
<b>Torch mounting</b>	The Torch mounting should be in preferably vertical design.	
<b>Detector</b>	Preferably Charge coupled type detector with 1 million pixel.	
<b>Atmospheric correction</b>	Vacuum based atmospheric correction facility.	
<b>Gas option</b>	Full PC controlled gas control for Plasma, Nebulizer, carrier or any other gas is used.	
<b>RF generator</b>	Free Running solid State type 27 MHz/ 40 MHz 700– 1600 W, efficiency should be 75% or better	
<b>Gas Consumption</b>	Argon with purity should be minimum 99.995%	
<b>Gas saver mode</b>	System should have gas saving mode without shutting down plasma.	
<b>Configuration</b>	Configuration	Dual view (Axial & radial) configuration to select axial or radial view with automatic switchover possibility and auto alignment.
	Torch	Aqueous quartz torch
	Torch mounting	Vertical mounting
	Spray chamber	Glass cyclone
	Nebulizer	Coaxial glass / glass concentric nebulizer
	Peristaltic pump	4 channels, 12-roller pump
<b>Hydride Kit</b>	Continuous hydride Generation assembly with peristaltic pump with External Gas Liquid Separator.	
<b>Software</b>	The PC controlled software to be quoted along with ICP system	
	Software should be window 10 based or latest version, Instrument operation and control, data handling and storage, report generation, calibration graph, results display and printing.	
	Software should be able to perform both qualitative and quantitative analysis	
	Software should be equipped with Built-in database/ suitable library for all wavelength.	
	Software should have calibration and Background correction option as standard feature.	
	Should have an option to develop method for unknown and complex matrix and petrochemical and crude oil.	
	Should be able check the suitable wavelength selection	
	Should have all QC/QA function and data export and printing option built-in with the software.	
	Should have all QC/QA function and data export and printing option built-in with the software.	
<b>Installation kit</b>	All installation accessory to be quoted along with system.	

<b>Aqueous sample handling kit</b>	The Aqueous sample handling kit should include.
	Aqueous Torch 2 Nos.
	Sample tubing – 20 Nos.
	Nebulizer – 1 Nos
	Spray Chamber – 1 Nos
<b>Exhaust Hood</b>	Compatible exhaust hood system
<b>Auto Sampler</b>	Auto-sampler for minimum 60 samples capacity with integrated pump & rinse station
<b>Installation kit</b>	Complete installation and test kit
<b>Warranty</b>	Fully 3 years Warranty & Additional service for two years.
<b>Gas Cylinder with Regulator</b>	Argon gas Cylinder– 04 nos. Gas Control Box
<b>UPS</b>	10 KVA Suitable UPS with 120mins back up
<b>Data Station</b>	Branded PC and LaserJet Printer
<b>Regulator</b>	2 Nos

## **8.PH/EC SOIL ANALYSER (SP 2000 SERIES)**

### **Fully automatic Robotic Analyzer with four probe measurement of pH and Conductivity in Soil Samples with integration of extraction procedure**

A combined one system for the measurement of pH & EC should include:

- Robotic handling system to make measurement of samples simultaneously using multiple brand of pH meters & EC electrodes.
- Complete automation functions for liquid handling, mixing, sample pipetting.
- Should Include 60 sample positions for 50m bottle to measure pH/ EC.
- System should be capable of analyzing more than 400 samples per day for both pH/EC.
- Possibility to upgrade the system to handle up to 352 samples per Hour.
- Robotic Arm Moving mechanism: XYZ with safety protection.
- Safety protection glass to be part of the system to avoid any manual intervene
- Simultaneous measurement of 4 samples at a time using probes for both pH/EC analysis.
- Should include addition of extraction solution, stirring as part of the analysis automatically.
- After settling time, the measurement to be done with automatic stirring
- Should have facility to upgrade in future to do capping/de-capping facility within the system for any kind of sample bottles
- Automated Liquid level and leak detection should be there with in the system.
- Rinse vessel, Cables and tubing, Multiple task facility, Instruction manual

- Locked protective front and side covers complying with CE (Council of European Union) regulations
  - Rinsing station for automatic rinsing of the probes, stirrers and needles in between samples
  - software controlled continuous refreshing of rinsing vessel to avoid cross-contamination between samples
- Power requirements: 220-240V
- at least 4 numbers of Additional sample trays and 3 sets of pH and EC electrodes need to be provided along with the instrument
  - Instrument quoted must be complete with all accessories which are required for installation like 2KVA UPS with 2 Hrs Back up suitable Work table with Granite Floor and calibration chemical need to be supplied by the vendor
  - Addition of bottles, trays & work list extension during run
  - A set of User manual ( both Soft and Hard Copy) need to be supplied by the vendor
  - Impart Training on Instrument (Technically and general maintenance), training should be provided at 2 periodic in an annum for a given set of users.
  - Latest Branded Computer with Laser Jet printer and Windows™ software
  - Warranty is for 3 years ( both consumables and non-consumables should be included during the warranty period)

### **SOFTWARE SPECIFICATIONS:**

- Definable access levels and password protection
- Fast set-up of sample work list in routine operation by dragging the racks to the analyzer and selecting the required application
- User definable sample table and print report layout
- Customized or pre-defined application files such as BOD (Biological Oxygen Demand), COD( Chemical Oxygen Demand), pH, Alkalinity, Turbidity, ISE(Ion Selective Electrode )etc.
- Direct data transfer to LIMS ( laboratory Information Management Systems) & spread sheets
- Possibility of exporting results during analysis
- Extensive Quality Control criteria including CLP (Contract Laboratory Program) protocols

- Audit Trail
- Possibility of using Quality samples and creating Quality Control Charts
- Automatic back-up of raw & calculated data meeting GLP ( Good Laboratory Practices) criteria
- Automatic flagging of “out-of-range” results applying pre-defined criteria
- Automatic sample result averaging applying pre-defined criteria
- Automatic start-up and shut-down using an analyses scheduler
- Overload alarm, filling failure, read-out stability check and continuous electrode control Calculations compliant with all (inter)national regulations.

**9. SET OF SOLAR BATTERIES (12V/200 AHS TUBLER BATTERIES) : WITH 4 YEARS WARRENTY**

**10. SOLAR PANELS 250 WTS CAPACITY:**  
( The offerer has to furnish the specifications of the equipment)

**11. INVERTERS 5 KVA:**  
( The offerer has to furnish the specifications of the equipment)



# **PRICE QUOTE**

**ANNEXURE- III**  
**PRICE QUOTE**

*(As per specifications mentioned in Annexure-II)*

**SCHEDULE**

Sl. No.	Description of equipment	Rate quoted per each		
		Rs.in figures (in words)		
		Basic price	Taxes	Total
1	PH METER ELECTRODES			
2	EC METER			
3	EC METER ELECTRODES			
4	AUTO SAMPLER TEST TUBES FOR MPAES-4200			
5	ACETYLENE GAS CYLINDER WITH REGULATER			
6	UV VISABLE SPECTRAPHOTOMETER WITH PC ATTACHMENT			
7	ICPOES(INDUCTIVELY COUPLED PLASMA OPTICAL EMISSION SPECTOMETRY)			
8	PH/EC SOIL ANAYLISER (SP 2000 SERIES)			
9	SET OF SOLAR BATTERIES (12V/200 AHS TUBLER BATTERIES)			
10	SOLAR PANELS 250 WTS CAPACITY			
11	INVERTERS 5KVA			

1. The offerer has to quote their lowest rate.
2. The basic rates quoted should be inclusive of all Taxes / Duties on FOL delivery to the concerned lab & inclusive of loading, unloading, training to concerned officials with machine operation etc. GST shall be shown separately.
3. The rates quoted in the Price quote should be firm and valid up to 31.03.2020. The APSAIDC Ltd. will not entertain any upward price revisions during the agreement period. The validity period can be extended on mutual consent of both the parties.
4. The rates and units shall not be over written. The rates and units shall always be in both figures and words.

Place :

Date

SIGNATURE OF OFFERER WITH SEAL

**ANNEXURE - IV**  
**DECLARATION ON Rs.100/- NON JUDICIAL STAMP PAPER**

1. I/We have read and understood the terms and conditions of EOI schedule relevant to “EOI notice” dated: 20-06-2019 and I/we have submitted the Annexures in accordance with the terms and conditions of above EOI schedule.
2. I/We shall not sell the products being quoted through this EOI schedule, at a lower price either to Government / Private parties in the State of Andhra Pradesh by way of giving any additional trade discount or incentives. If any such incident is noticed the A.P.S.A.I.D.C. Ltd. has the right to restrict all payments to such a lower rate apart from initiating penal action.
3. The information furnished in the Annexure-I is true and factual I clearly understand that our EOI schedule is liable for rejection, if any information furnished is found not true and factual, at any point of time.
4. The product/s offered by us is not copied or duplicated from any product design. In case the Corporation authorities find at any Time our products are an imitation or duplicate, we are bound for any legal action to be taken on us including blacklisting our firm.
5. I/We declare that I/we do not violate any patent/trade mark rights of any other manufacturer.

Place :  
Date :

SIGNATURE OF THE OFFERER  
WITH SEAL

## DATA SHEET

### Information to EOI applicant(s)

Clause Ref.	Description								
1	<p><b>The name of the Client is :</b></p> <p>Vice Chairman &amp; Managing Director                      APS Agro Industries Development Corporation Limited                      Regd. Office: No.55-17-224, 5<sup>th</sup> floor, Stalin Corporate, 2<sup>nd</sup> cross road                      Industrial Estate, Autonagar, Vijayawada-520007</p> <p>Ph: 0866 – 2556803 : Fax : 0866 - 2556804                      Email: <a href="mailto:apagrosdmae@gmail.com">apagrosdmae@gmail.com</a>, Website: <a href="http://www.apagros.org">www.apagros.org</a></p>								
2	<p><b>The Method of selection is :</b></p> <p>Quality-and Cost-Based Selection (QCBS)</p>								
3	<p><b>The Edition of the Guidelines is :</b></p> <p>As furnished in the EOI schedule</p>								
4	<p><b>EOI and Financial Schedules are requested Separately:</b>                      Yes</p>								
5	<p><b>Particulars of materials to be supplied:</b>                      Supply, installation, commissioning, validation of equipment to Soil Testing Labs.</p>								
6	<p><b>The addresses and Telephone numbers of client's Regional offices are</b></p>								
	<table border="1" style="width: 100%;"> <tr> <td style="text-align: center;">1</td> <td>                     District Manager.                      APS AGRO IND. DEV. CORP. LTD.                      Thotapalem road, Opp. RTC Bus stand,                      Vizianagaram 531202, Andhra Pradesh,                      Phone No: 08922-224256. Mobile No. <b>9133460123</b> </td> </tr> <tr> <td style="text-align: center;">2</td> <td>                     Manager(AE),                      APS AGRO IND. DEV. CORP. LTD.                      Pension office lane, Maidan grounds,                      Eluru 534 005, Andhra Pradesh,                      phone No. 08812-230731 Mobile: <b>9848076866</b> </td> </tr> <tr> <td style="text-align: center;">3</td> <td>                     Regional Manager                      APS AGRO IND. DEV. CORP. LTD.                      Besides IMFL Godown, Nallapadu                      Guntur 522006, Andhra Pradesh                      Phone No.0863-2290299, Mobile: <b>9948099876</b> </td> </tr> <tr> <td style="text-align: center;">5</td> <td>                     Regional Manager,                      APS AGRO IND. DEV. CORP. LTD.                      Industrial Estate, Bellary road,                      Anantapur- 515 001, Andhra Pradesh.                      Phone No.08554-270815, Mobile: <b>9948076847</b> </td> </tr> </table>	1	District Manager. APS AGRO IND. DEV. CORP. LTD. Thotapalem road, Opp. RTC Bus stand, Vizianagaram 531202, Andhra Pradesh, Phone No: 08922-224256. Mobile No. <b>9133460123</b>	2	Manager(AE), APS AGRO IND. DEV. CORP. LTD. Pension office lane, Maidan grounds, Eluru 534 005, Andhra Pradesh, phone No. 08812-230731 Mobile: <b>9848076866</b>	3	Regional Manager APS AGRO IND. DEV. CORP. LTD. Besides IMFL Godown, Nallapadu Guntur 522006, Andhra Pradesh Phone No.0863-2290299, Mobile: <b>9948099876</b>	5	Regional Manager, APS AGRO IND. DEV. CORP. LTD. Industrial Estate, Bellary road, Anantapur- 515 001, Andhra Pradesh. Phone No.08554-270815, Mobile: <b>9948076847</b>
1	District Manager. APS AGRO IND. DEV. CORP. LTD. Thotapalem road, Opp. RTC Bus stand, Vizianagaram 531202, Andhra Pradesh, Phone No: 08922-224256. Mobile No. <b>9133460123</b>								
2	Manager(AE), APS AGRO IND. DEV. CORP. LTD. Pension office lane, Maidan grounds, Eluru 534 005, Andhra Pradesh, phone No. 08812-230731 Mobile: <b>9848076866</b>								
3	Regional Manager APS AGRO IND. DEV. CORP. LTD. Besides IMFL Godown, Nallapadu Guntur 522006, Andhra Pradesh Phone No.0863-2290299, Mobile: <b>9948099876</b>								
5	Regional Manager, APS AGRO IND. DEV. CORP. LTD. Industrial Estate, Bellary road, Anantapur- 515 001, Andhra Pradesh. Phone No.08554-270815, Mobile: <b>9948076847</b>								

6	Regional Manager APS AGRO IND. DEV. CORPN. LTD. Vellore Road Chittoor- Andhra Pradesh Mobile : <b>9948076847</b>
7	<b>The client will provide the following:</b> Technical specifications of the Soil Testing lab equipment,.
8	<b>The EOI submission address is:</b> Vice Chairman & Managing Director APS Agro Industries Development Corporation Limited Regd. Office: No.55-17-224, 5 <sup>th</sup> floor, Stalin Corporate, 2 <sup>nd</sup> cross road Industrial Estate, Autonagar, Vijayawada-520007
9	The information on the outer envelope should also include: <b>EOI for Supply and installation, commissioning, validation of equipment to Soil Testing lab equipment.</b> Validity of EOI shall be 90 days from the date of submission.